

CALIBRATION WORK SHEET pre check

Date of Calibration: 4/18/2015
 Technician: Gelpi

Sonde ID: 0360d61AA

RP DO membrane changed? Y *Note: Wait 3 to 6 hours before calibrating for unattended*
 RP DO membrane o-ring changed? Y *deployments; run in Discrete mode for 10 minutes to accelerate*
burn in. (Rapid Pulse DO Only)
 Turbidity wiper changed? ~~Y~~ N Chlorophyll wiper changed? Y N
 ROX DO wiper changed? ~~Y~~ N BGA-PE wiper changed? Y N
 BGA-PC wiper changed? ~~Y~~ N Rhodamine wiper changed? Y N

Note: If parking problems occur with optical probes having a serial number 07L (Dec 07) or older, be sure the firmware is 3.06 or later. Parking issues with optical probes having a serial number prior to 07L may be related to a dirty wiper body or pad.

Record sonde battery voltage: 12 (if applicable)

Record Calibration Values
 Standard Pre Cal / Post Cal

Record the following diagnostic numbers after calibration.

6560 Conductivity cell constant 5.32118 Range 5.0 ± .45
1-Cont. 5.32118
 Integrated conductivity cell constant _____ Range 5.0 ± .70
 pH mv Buffer 7 _____ Range 0 ± 50 mv
 pH mv Buffer 4 152.2 Range +180 ± 50 mv*
 pH mv Buffer 10 _____ Range -180 ± 50 mv*
 *Note: Millivolt span between pH 4 and 7 should be ≈ 165 to 180 mv
 Millivolt span between pH 7 and 10 should be ≈ 165 to 180 mv
 DO charge (RP only) 43.1 Range 25 to 75
 DO gain 1.18096 Range 0.7 to 1.4
 ODO gain _____ Range 0.85 to 1.15

Temperature	_____	Sonde
Conductivity	_____	/
pH 7	_____	/
pH 4	<u>4.0</u>	<u>4.03 378</u>
pH 10	_____	/
ORP	_____	/
Turbidity	_____	/
Turbidity	_____	/
Turbidity 0.5	_____	/
Chlorophyll	_____	/
Chlorophyll	_____	/
DO RP	_____	<u>98 / 100?</u>
DO ROX	_____	/
BGA PE/PC	_____	/
BGA PE/PC	_____	/
Rhodamine	_____	/

Turbidity standard used in calibration _____

Manufacturer and part number _____

25.4 x 1.45 = mult

Barometric Pressure: 752.714 mmHg

DO % Calculated - (BARO mmHg divided by 7.6) = % saturation

Example: 760 ÷ 7.6 = 100.0%

Depth Calibration - If zero was entered, record barometric pressure at time of calibration _____ mmHg

Depth Calibration - If offset depth was entered, record value _____ meters/feet and pressure _____ mmHg

Depth Calibration (Vented) - Acceptable calibration constant: 0.0 psig ± 0.15 _____

Notes:

*{ 432 tan 17:46:51 a check out before delivery to boat. pH was fine & measured 4.03
 { Phm 17:44 the solution of 4.0 that it was noted is.
 before assembly Conductivity value was erratic. First, measured 50 mS/cm at 48.
 I calibrated it to 50 then ran 10 mS/cm soln. That was off. Then
 calibrated 10 mS/cm. Then test 50 mS/cm again. Found 2 calibrations again*

$(pH-7) \times K = -920704$

$(pH-7) \times K / mv = -5.23737$

shall check conductivity measurement