

CALIBRATION WORK SHEET

Date of Calibration: _____
Technician: _____

Sonde ID: _____

RP DO membrane changed? Y N Note: Wait 3 to 6 hours before calibrating for unattended
RP DO membrane o-ring changed? Y N deployments; run in Discrete mode for 10 minutes to accelerate
burn in. (Rapid Pulse DO Only)
Turbidity wiper changed? Y N Chlorophyll wiper changed? Y N
ROX DO wiper changed? Y N BGA-PE wiper changed? Y N
BGA-PC wiper changed? Y N Rhodamine wiper changed? Y N

Note: If parking problems occur with optical probes having a serial number 07L (Dec 07) or older, be sure the firmware is 3.06 or later. Parking issues with optical probes having a serial number prior to 07L may be related to a dirty wiper body or pad.

Record sonde battery voltage: _____ (if applicable)

Record Calibration Values
Standard Pre Cal / Post Cal

Record the following diagnostic numbers after calibration.

6560 Conductivity cell constant _____ Range 5.0 ± .45

Temperature _____ Sonde _____

Integrated conductivity cell constant _____ Range 5.0 ± .70

Conductivity _____ / _____

pH mv Buffer 7 _____ Range 0 ± 50 mv

pH 7 _____ / _____

pH mv Buffer 4 _____ Range +180 ± 50 mv*

pH 4 _____ / _____

pH mv Buffer 10 _____ Range -180 ± 50 mv *

pH 10 _____ / _____

*Note: Millivolt span between pH 4 and 7 should be ≈ 165 to 180 mv

ORP _____ / _____

Millivolt span between pH 7 and 10 should be ≈ 165 to 180 mv

Turbidity _____ / _____

DO charge (RP only) _____ Range 25 to 75

Turbidity _____ / _____

DO gain _____ Range 0.7 to 1.4

Turbidity 0.5 _____ / _____

ODO gain _____ Range 0.85 to 1.15

Chlorophyll _____ / _____

Turbidity standard used in calibration _____

Chlorophyll _____ / _____

Manufacturer and part number _____

DO RP _____ / _____

Barometric Pressure: _____ mmHg

DO ROX _____ / _____

DO % Calculated – (BARO mmHg divided by 7.6) = % saturation

BGA PE/PC _____ / _____

Example: $760 \div 7.6 = 100.0\%$

BGA PE/PC _____ / _____

Rhodamine _____ / _____

Depth Calibration - If zero was entered, record barometric pressure at time of calibration _____ mmHg

Depth Calibration - If offset depth was entered, record value _____ meters/feet and pressure _____ mmHg

Depth Calibration (Vented) – Acceptable calibration constant: 0.0 psig ± 0.15 _____

Notes: _____